


Thoracolumbar Disease, Pain or Injury – Acute

Intervertebral disc disease (IVDD), back pain, blown disc, hemilaminectomy, pain from spondylosis, pain from kyphosis, pain from scoliosis, neuropathy, diffuse idiopathic skeletal hyperostosis, fibrocartilaginous embolism (FCE).



1. A small or large Assisi Loop can be used to treat the thoracolumbar (middle of the back) area in the pet depending on the size of the pet. Dachshund- and Chihuahua-sized pets can use a small Assisi Loop. Large-breed dogs need the larger Loop. Positioning is extremely important for optimum effect.
2. Position the pet lying down either flat on the ground or up on their elbows.
3. The thoracolumbar area can be felt by feeling the end of the ribs, following them up to the **thoracolumbar junction (arrow)** and placing the center of the large Assisi Loop at that point. **See Pictures and Diagrams**
4. The Loop can also be placed around the patient if the size of the patient and the size of the Assisi Loop are appropriate for this positioning to be utilized. This type of positioning allows the largest area of the spine to be treated and should be utilized if possible. To place the Loop around the thoracolumbar area of the pet, it is usually easier to place the Assisi Loop around the rear feet and slide it up over the hips, making sure the tail goes through the Loop as it is being slid up the rear limbs. The Assisi Loop is slid to the area where the end of the ribs is felt along the back. **See Pictures and Diagrams**
5. **CAUTION: Do not force the pet's shoulders, haunches, or torso through the coil. If they do not slip easily through the coil, place the coil flat against, over, or under animal's torso, as shown in the related images. Forcing the animal through the coil can be traumatic to the pet and may overly stress the connection of the coil to the circuit board inside the Loop case.**
6. If the pet will not lie down, an Assisi LOOP-AID can be used. Attach the Assisi Loop with two Velcro straps to maintain the Loop position with the center of the Assisi Loop over the thoracolumbar joint. **See Pictures**
7. The signal generator can be positioned to help balance and stabilize the unit during treatment.
8. The signal generator should be located in view of the person treating the pet.

9. Do not leave the pet unattended when using this placement as small movements can change the positioning of the Loop and the area being treated.
10. Turn the unit on by depressing the button on the front of the white rubberized signal generator; green lights will illuminate about once per second. If the light is blinking 2-3 times per second, the battery is running low and a new Assisi Loop should be purchased.
11. Keep the unit in place for the full 15 minutes until the unit shuts off.
12. Use the unit 4 times daily, if possible, for 5 days. Treat twice daily for an additional 5 days or until pain is completely resolved.
13. There should be a minimum of 2 hours between treatments for optimum battery efficiency.
14. If Prednisone is administered and tapered, the Assisi Loop can be tapered after the Prednisone is discontinued.
15. Thoracolumbar disease can be a lifelong condition where significant activity creates inflammation. Many patients benefit significantly from having once-daily treatments after the clinical signs have resolved to prevent recurrences.



Pet Owner Responsibility. A panel of licensed, experienced veterinarians has developed this Guide. Pet owners should use this protocol only to treat a condition listed above and as directed by a qualified veterinary clinician. Other protocols in this Guide should be used only as directed by a qualified veterinary clinician.

Important: Before starting treatment, please consult product label for optimal product use and safety guidelines. To learn more about the technology, visit our website.